

PACKAGING

CK-NAC / CK-MB Control Level 1	Ref.: 101-0697	Cont.: 2 mL
CK-NAC / CK-MB Control Level 2	Ref.: 101-0762	Cont.: 2 mL

Store at 2 - 8°C

PRODUCT CHARACTERISTICS

CK-Nac / CK-MB CONTROL is a human lyophilised control serum.
Is intended for control of accuracy for use with manual and automated analytical procedures.

REAGENTS

Human serum. Biological additives. Bacteriostatics agents.
The concentration of the components are lot-specific.
See chart below for the values of ingredients.

PRECAUTIONS

Components from human origin have been tested and found negative for the presence of HBsAg, HCV and antibody to HIV (1/2). However handle cautiously as potentially infectious.

PREPARATION

-Reconstitute (→) with 2.0 mL of distilled water.
-Mix thoroughly, avoiding foam forming.
-Bring to room temperature for about 30 min. before use.
Improper handling and/or storage can affect results.
Inaccurate reconstitution and errors in assay technique can cause erroneous results.

STORAGE AND STABILITY

The Control serum is stable until the expiration date on the label when stored tightly closed at 2-8° C and contaminations are prevented during their use. Do not use reagents over the expiration date or if there is visible evidence of microbial grown.

Store tightly capped when not use.

- After reconstitution is stable for:
- At 15° C to 25° C: 8 hours
- At 2° C to 8° C: 5 days
- At -25° C to -15° C: 1 month

LOT			
------------	--	---	--

COMPONENT	METHOD	T ^a	VALUE	RANGE	UNITS
CK-Nac	DGKC	37° C			U / L
CK-MB	Immunological method	37° C			U / L

*This Control data sheet for lot is applicable to sublots. Sequential alphabetical letter (e.g. A, B, C etc.) following the lot number.