

PACKAGING

Ref.: 101-0241	Cont.: 5 x 20 mL
Ref.: 101-0016	Cont.: 4 x 50 mL
Ref.: 101-0052	Cont.: 6 x 100 mL
Ref.: 101-0053	Cont.: 4 x 250 mL
Ref.: 101-0268	Cont.: 12 x 50 mL



Store at 2 - 8° C

CLINICAL SIGNIFICANCE

Triglycerides are fats that provide energy for the cell. Like cholesterol, they are delivered to the body's cells by lipoproteins in the blood. A diet with a lot of saturated fats or carbohydrates will raise the triglyceride levels. The increases in serum triglycerides are relatively non-specific. For example liver dysfunction resulting from hepatitis, extra hepatic biliary obstruction or cirrhosis, diabetes mellitus is associated with the increase^{3,6,7}. Clinical diagnosis should not be made on a single test result; it should integrate clinical and other laboratory data.

PRINCIPLE OF THE METHOD

Sample triglycerides incubated with lipoproteinlipase (LPL), liberate glycerol and free fatty acids. Glycerol is converted to glycerol-3-phosphate (G3P) and adenosine-5-diphosphate (ADP) by glycerol kinase and ATP. Glycerol-3-phosphate (G3P) is then converted by glycerol phosphate dehydrogenase (GPO) to dihydroxyacetone phosphate (DAP) and hydrogen peroxide (H₂O₂). In the last reaction, hydrogen peroxide (H₂O₂) reacts with 4-aminophenazone (4-AP) and p-chlorophenol in presence of peroxidase (POD) to give a red colored dye:



The intensity of the color formed is proportional to the triglycerides concentration in the sample^{1,2,3}.

REAGENTS

R 1	GOOD pH 7.5	50 mmol/L
Buffer	p-Chlorophenol	2 mmol/L
R 2	Lipoprotein lipase (LPL)	150000 U/L
	Glycerolkinase (GK)	500 U/L
	Glycerol-3-oxidasa (GPO)	2500 U/L
	Peroxidase (POD)	440 U/L
	4 - Aminophenazone (4-AP)	0.1 mmol/L
Enzymes	ATP	0.1 mmol/L
TRIGL. CAL	Triglycerides aqueous primary standard 200 mg/dL	

Optional (not included in the kit)

Contro-N	Ref.: 101-0252	4 x 5 mL	Lyophilized human control serum
	Ref.: 101-0083	20 x 5 mL	
Contro-P	Ref.: 101-0253	4 x 5 mL	Lyophilized human control serum
	Ref.: 101-0084	20 x 5 mL	

PREPARATION

Working reagent (WR): Dissolve (→) the contents of one vial R 2 Enzymes into one bottle of R 1 Buffer.

Ref: 1001310 Working reagent (WR): Dissolve (→) the contents of one vial R 2 Enzymes in 10 mL of R 1 Buffer.

Cap and mix gently to dissolve contents.

WR stability: 6 weeks at 2 - 8° C or 1 week at room temperature (15 - 25° C).

STORAGE AND STABILITY

All the components of the kit are stable until the expiration date on the label when stored tightly closed at 2 - 8° C, protected from light and contaminations prevented during their use. Do not use reagents over the expiration date.

Signs of reagent deterioration:

- Presence of particles and turbidity.
- Blank absorbance (A) at 505 nm ≥ 0.14.

ADDITIONAL EQUIPMENT

- Spectrophotometer or colorimeter measuring at 505 nm.
- Matched cuvettes 1.0 cm light path.
- General laboratory equipment.

SAMPLES

Serum or heparinized or EDTA plasma¹. Stability of the sample: 5 days at 2 - 8° C .

PROCEDURE

Notes: CHRONOLAB SYSTEMS has instruction sheets for several automatic analyzers. Instructions for many of them are available on request.

TRIGLYCERIDES CAL: Proceed carefully with this product because due its nature it can get contaminated easily.

LCF (Lipid Clearing Factor) is integrated in the reagent.

Calibration with the aqueous Standard may cause a systematic error in automatic procedures. In these cases, it is recommended to use a serum Calibrator.

Use clean disposable pipette tips for its dispensation.

- Assay conditions:
Wavelength: 505 nm (490 - 550)
Cuvette: 1 cm light path
Temperature 37° C / 15 - 25° C
- Adjust the instrument to zero with distilled water.
- Pipette into a cuvette:

	Blank	Standard	Sample
WR (mL)	1.0	1.0	1.0
Standard ^(Note 1,2) (µL)	--	10	--
Sample (µL)	--	--	10

- Mix and incubate for 5 min. at 37° C or 10 min. at room temperature.
- Read the absorbance (A) of the samples and Standard, against the Blank. The colour is stable for at least 30 minutes.

CALCULATIONS

$$\frac{(A)\text{Sample}}{(A)\text{Standard}} \times 200 \text{ (Standard conc.)} = \text{mg/dL triglycerides in the sample}$$

Conversion factor: mg/dL x 0.0113= mmol/L.

QUALITY CONTROL

Control sera are recommended to monitor the performance of assay procedures.

If control values are found outside the defined range, check the instrument, reagents and calibrator for problems.

Each laboratory should establish its own Quality Control scheme and corrective actions if controls do not meet the acceptable tolerances.

REFERENCE VALUES

Men	40 – 160 mg/dL
Women	35 – 135 mg/dL

These values are for orientation purpose; each laboratory should establish its own reference range.

PERFORMANCE CHARACTERISTICS

Measuring range: From detection limit of 0.000 mg/dL to linearity limit of 2200 mg/dL.

If the results obtained were greater than linearity limit, dilute the sample 1/2 with NaCl (9 g/L) and multiply the result by 2.

Precision:

	Intra-assay (n=20)		Inter-assay (n=20)	
Mean (mg/dL)	103	219	103	217
SD	0.41	0.93	3.74	7.80
CV (%)	0.39	0.43	3.62	3.59

Sensitivity: 1 mg/dL = 0.00137 A.

Accuracy: Results obtained using CHRONOLAB reagents (y) did not show systematic differences when compared with other commercial reagents (x).

The results obtained using 50 samples were the following:

Correlation coefficient (r): 0.99760.

Regression equation: $y = 0.905x + 10.8$.

The results of the performance characteristics depend on the analyzer used.

INTERFERENCES

No interferences were observed with bilirubin up to 170 $\mu\text{mol/L}$ and hemoglobin up to 10 g/L^2 .

A list of drugs and other interfering substances with cholesterol determination has been reported by Young et al ^{4,5}.

BIBLIOGRAPHY

1. Buccolo G et al. Quantitative determination of serum triglycerides by use of enzymes. Clin Chem 1973; 19 (5): 476-482.
2. Fossati P et al. Clin. Chem 1982; 28(10): 2077-2080.
3. Kaplan A et al. Tryglycerides. Clin Chem The C.V. Mosby Co. St Louis. Toronto. Princeton 1984; 437 and Lipids 1194-1206.
4. Young DS. Effects of drugs on Clinical Lab. Tests, 4th ed AACC Press, 1995.
5. Young DS. Effects of disease on Clinical Lab. Tests, 4th ed AACC 2001.
6. Burtis A et al. Tietz Textbook of Clinical Chemistry, 3rd ed AACC 1999.
7. Tietz N W et al. Clinical Guide to Laboratory Tests, 3rd ed AACC 1995.