

PACKAGING

Ref.: 101-0378	Cont.: 19 x 3 mL
----------------	------------------

Store at 2-8°C

CLINICAL SIGNIFICANCE

Glucose forms stable glycosylated serum proteins with several plasmatic proteins, mainly, albumin, in covalent union. The determination of fructosamina is based on the measurement of these glycoproteins.

The measurement of fructosamine has utility to know retrospectively (2-3 weeks) the level of glucose concentration in blood.

This test is used for control and monitoring of diabetic patients and not for diagnosis^{5,6}.

Clinical diagnosis should not be made on a single test result; it should integrate clinical and other laboratory data.

PRINCIPLE OF THE METHOD

Under alkaline conditions the fructosamine or glycosylated serum proteins reduce a nitro-blue tetrazolium chloride (NBT) salt.

The colour developed is directly proportional to the serum fructosamine concentration¹.

REAGENTS

R 1 Buffer	Carbonate Detergents	200 mmol/L
R 2 Enzymes	Nitrotetrazolium chloride (NBT) Uricase	0.48 mmol/L 3000 U/L
FRUCTOSAMINE CAL	Calibrator lyophilised serum	

Optional (not included in the kit)

Contro-N	Ref.: 101-0252	4 x 5 mL	Lyophilized human control serum
	Ref.: 101-0083	20 x 5 mL	
Contro-P	Ref.: 101-0253	4 x 5 mL	Lyophilized human control serum
	Ref.: 101-0084	20 x 5 mL	

PREPARATION

- Working reagent (WR):

Dissolve (→) 1 tablet of R 2 Enzymes with one vial R 1 Buffer.
Cap vial and mix gently to dissolve contents.
The reagent is stable after reconstitution 15 days at 2-8° C or 5 days at room temperature (15-25° C). Protected from light.

- FRUCTOSAMINE CAL:

Dissolve (→) the contents of one vial Calibrator with 1 mL of distilled water. Cap vial and mix gently to dissolve contents.
The reconstituted calibrator is stable 15 days at 2-8° C or 2 months at -20° C.

STORAGE AND STABILITY

All the components of the kit are stable until the expiration date on the label when stored tightly closed at 2-8° C, protected from light and contaminations prevented during their use.

Do not use the tablets if appears broken.

Do not use reagents over the expiration date.

Signs of reagent deterioration:

- Presence of particles and turbidity.
- Blank absorbance (A) at 520 ≥ 0.30.

ADDITIONAL EQUIPMENT

- Spectrophotometer or colorimeter measuring at 520 nm.
- Matched cuvettes 1.0 cm light path.
- General laboratory equipment.

SAMPLES

Serum^{1,2}:

Don't use haemolized samples. Separated from cells as rapidly as possible.
Stability of the sample: 7 days at 2-8° C.

PROCEDURE

Notes: CHRONOLAB SYSTEMS has instruction sheets for several automatic analyzers. Instructions for many of them are available on request.

- Assay conditions:
Wavelength: 520 (490-550) nm
Cuvette: 1 cm light path
Temperature: 37° C
- Adjust the instrument to zero with distilled water.
- Pipette into a cuvette:

	Blank	Calibrator	Sample
WR (mL)	1.0	1.0	1.0
Calibrator (µL)	--	100	--
Sample (µL)	--	--	100

- Mix, incubate at 37° C and start stopwatch.
- Read the absorbance (A₁) of the calibrator and sample exactly after 10 min and after 15 min (A₂) of the sample addition against distilled water.
- Calculate: $\Delta A = A_2 - A_1$.

CALCULATIONS

$$\frac{(\Delta A)_{\text{Sample}}}{(\Delta A)_{\text{Calibrator}}} \times \text{Calibrator conc.} = \mu\text{mol/L of fructosamine in the sample}$$

QUALITY CONTROL

Control sera are recommended to monitor the performance of assay procedures.

If control values are found outside the defined range, check the instrument, reagents and calibrator for problems.

Each laboratory should establish its own Quality Control scheme and corrective actions if controls do not meet the acceptable tolerances.

REFERENCE VALUES¹

In nondiabetic samples: 187 – 287 µmol/L¹

These values are for orientation purpose; each laboratory should establish its own reference range.

PERFORMANCE CHARACTERISTICS

Measuring range: From *detection limit* of 1 µmol/L to *linearity limit* of 1000 µmol/L.

If the results obtained were greater than linearity limit, dilute the sample 1/2 with NaCl (9 g/L) and multiply the result by 2.

Precision:

	Intra-assay (n=20)		Inter-assay (n=20)	
	Mean (µmol/L)	217	587	197
SD	4.71	8.31	4.20	10.2
CV (%)	2.17	1.41	2.12	1.85

Sensitivity: 1 µmol/L = 00020 A.

Accuracy: Results obtained using CHRONOLAB reagents (y) did not show systematic differences when compared with other commercial reagents (x).

The results obtained using 50 samples were the following:

Correlation coefficient (r): 0.992

Regression equation: $y=0.991x + 1.473$

The results of the performance characteristics depend on the analyzer used.

INTERFERENCES

Do not interfere: Hemolysis up to 5 g/L, bilirubin up to 20 mg/dL and triglycerides up to 6 gr/L^{1,2}.

A list of drugs and other interfering substances with fructosamine determination has been reported by Young et. al^{3,4}.

BIBLIOGRAPHY

1. Baker JR. et al. Use of protein-based standards in automated colorimetric determinations of fructosamine in serum. Clin Chem 1985; (31/9): 1550-1554.
2. Hurst P. Clin Chem 1987; (33/10): 1947.
3. Young DS. Effects of drugs on Clinical Lab. Tests, 4th ed AACC Press, 1995.
4. Young DS. Effects of disease on Clinical Lab. Tests, 4th ed AACC 2001.
5. Burtis A et al. Tietz Textbook of Clinical Chemistry, 3rd ed AACC 1999.
6. Tietz N W et al. Clinical Guide to Laboratory Tests, 3rd ed AACC 1995.