



PACKAGING

Calibrator CFAS	Ref.: 101-0353	Cont.: 20 x 3 mL
Calibrator CFAS	Ref.: 101-0354	Cont.: 4 x 3 mL

Store at 2 - 8°C.

PRODUCT CHARACTERISTICS

CALIBRATOR CFAS is a human lyophilised serum.

The concentrations and activities have been selected so as to ensure optimum calibration for use in manual and automatic analysers.

REAGENTS

Human serum. Biological additives. Bacteriostatics agents.

The concentration / activities of the components are lot-specific. The exact values and ranges valid for reagents are given informational purpose in the value sheet.

PRECAUTIONS

The individual units comprising lot or lots of components from human origin, have been tested for HBsAg and antibodies to HIV and found negative / non reactive by an FDA approved test. However, handle cautiously as potentially infectious.

CALIBRATION

Information available on request.

PREPARATION

-Reconstitute (→) with 3.0 mL of distilled water.

-Mix thoroughly, avoiding foam forming.

-Bring to room temperature for about 30 min. before use.

Improper handling and/or storage can affect results.

Inaccurate reconstitution and errors in assay technique can cause erroneous results.

STORAGE AND STABILITY

The calibrator is stable until the expiration date on the label when stored tightly closed at 2-8° C and contaminations are prevented during their use. Do not use reagents over the expiration date or if there is visible evidence of microbial grown. Store calibrators tightly capped when not use.

- After reconstitution is stable for:

At 15° C to 25° C: 8 hours

At 2° C to 8° C: 2 days

At -25° C to -15° C: 4 weeks

- Bilirubin (stored protected from light):

At 15° C to 25° C: 6 - 3 hours (total-direct)

At 2° C to 8° C: 24 - 8 hours (total-direct)

At -25° C to -15° C: 2 weeks

- Acid phosphatase:

At 15° C to 25° C: 4 hours

At 2° C to 8° C: 1 day

At -25° C to -15° C: 2 weeks

PIQCC148e
V 2018/5

CHRONOLAB SYSTEMS, S.L., Travessia Prat de la Riba 34 B,
08849 Sant Climent de Llobregat, Barcelona, Spain
Tel. +34 617722466, www.chronolab.com, e-mail: info@chronolab.com



PACKAGING

Calibrator CFAS	Ref.: 101-0353	Cont.: 20 x 3 mL
Calibrator CFAS	Ref.: 101-0354	Cont.: 4 x 3 mL

Store at 2 - 8°C.

PRODUCT CHARACTERISTICS

CALIBRATOR CFAS is a human lyophilised serum.

The concentrations and activities have been selected so as to ensure optimum calibration for use in manual and automatic analysers.

REAGENTS

Human serum. Biological additives. Bacteriostatics agents.

The concentration / activities of the components are lot-specific. The exact values and ranges valid for reagents are given informational purpose in the value sheet.

PRECAUTIONS

The individual units comprising lot or lots of components from human origin, have been tested for HBsAg and antibodies to HIV and found negative / non reactive by an FDA approved test. However, handle cautiously as potentially infectious.

CALIBRATION

Information available on request.

PREPARATION

-Reconstitute (→) with 3.0 mL of distilled water.

-Mix thoroughly, avoiding foam forming.

-Bring to room temperature for about 30 min. before use.

Improper handling and/or storage can affect results.

Inaccurate reconstitution and errors in assay technique can cause erroneous results.

STORAGE AND STABILITY

The calibrator is stable until the expiration date on the label when stored tightly closed at 2-8° C and contaminations are prevented during their use. Do not use reagents over the expiration date or if there is visible evidence of microbial grown. Store calibrators tightly capped when not use.

- After reconstitution is stable for:

At 15° C to 25° C: 8 hours

At 2° C to 8° C: 2 days

At -25° C to -15° C: 4 weeks

- Bilirubin (stored protected from light):

At 15° C to 25° C: 6 - 3 hours (total-direct)

At 2° C to 8° C: 24 - 8 hours (total-direct)

At -25° C to -15° C: 2 weeks

- Acid phosphatase:

At 15° C to 25° C: 4 hours

At 2° C to 8° C: 1 day

At -25° C to -15° C: 2 weeks

PIQCC148e
V 2018/5

CHRONOLAB SYSTEMS, S.L., Travessia Prat de la Riba 34 B,
08849 Sant Climent de Llobregat, Barcelona, Spain
Tel. +34 617722466, www.chronolab.com, e-mail: info@chronolab.com